



Accreditation Certificate No. (020906/1A)

**Arab Republic of Egypt
Egyptian Accreditation Council (EGAC)**

**Certifies that
Quality Control Laboratory
ATOS Pharma Company - SEKEM**

**El Katiba – Belbies City
Al Sharqia Governorate - Egypt**

**Has been accredited by EGAC in compliance with the requirements of
ISO/IEC 17025:2017**

**In Some Microbiological Tests for Food, Herbs and
Water and Some Chemical Tests for Herbs, Spices,
Water, Honey and Some Pharmaceutical Products**

**The scope of accreditation is described in the attached schedule No. (020906/1B)
Scope Issue No. (07)**

Issue No. (07): February 19, 2026

Valid to: December 23, 2029

Subject to continued compliance to the above standard and EGAC requirements

The Laboratory is accredited to issue reports / certificates under EGAC
accreditation according to the attached scope of accreditation

EGAC is an ILAC MRA Signatory in the Field of Calibration, Testing, Medical Labs,
Proficiency Testing Providers and Inspection Bodies Accreditation

1st Accreditation Date: December 24, 2009

Mohamed Hassan Abd El Halim

Mohamed Hassan

Executive Director

Egyptian Accreditation Council



73837

شهادة اعتماد رقم (أ١/٠٢٠٩٠٦)



جمهورية مصر العربية المجلس الوطني للاعتماد (إيجاك)

يشهد بأن
معمل ضبط الجودة

شركة اتوس فارما - سيكم

الكتيبة - مدينة بابيس

محافظة الشرقية - مصر

قد حصل على اعتماد المجلس حيث تم تقييمه طبقاً للمواصفة الدولية
ISO/IEC 17025:2017

في بعض الاختبارات الميكروبيولوجية للأغذية والاعشاب والمياه وبعض الاختبارات
الكيميائية للتوابل والاعشاب والعسل والمياه وبعض المستحضرات الطبية

الموضحة بالبيان المرفق رقم (أ١/٠٢٠٩٠٦)

رقم إصدار المجال (٠٧)

اصدار رقم (٠٧): ١٩ فبراير ٢٠٢٦ صالحة حتى: ٢٣ ديسمبر ٢٠٢٩

شرط الحفاظ على التوافق مع المواصفات أعلاه والمتطلبات الخاصة بالمجلس
المعمل له الحق في إصدار تقارير/ شهادات تحت مظلة المجلس الوطني للاعتماد
طبقاً لمجال الاعتماد المرفق

المجلس الوطني للاعتماد موقع على اتفاقية الاعتراف المتبادل مع منظمة التعاون الدولي
لاعتماد المعامل (ILAC) في مجال اعتماد معامل المعايرة والاختبار والتحليل الطبية
ومقدمي خدمة اختبارات الكفاءة الفنية وجهات التفتيش

تاريخ الاعتماد الأول: ٢٤ ديسمبر ٢٠٠٩

محمد حسن عبد الحليم

.....
المدير التنفيذي

للمجلس الوطني للاعتماد



73838

Schedule of Accreditation
for Testing Laboratory According to ISO/IEC 17025
Issued To
(Quality Control Laboratory)
(Atos Pharma Company – Sekem)
(Elkatiba – Belbies City – Sharkeia Government - Egypt)

Schedule No.:020906B

1st Accreditation Date:24-Dec-2009

Issue No. (06):24-Dec 2025

- Revision No. (0):

Valid to:23-Dec-2029

Materials / Products Tested	Types of Tests / Properties Measured / Range of Measurements	Standard Specifications / Techniques Used
Food and Animal feeding Stuff	Enumeration of P-glucuronidase -positive Escherich ia coli	ISO 16649-2:2001 Balance Model: Vibra AJ 2200 E S.N: BLI41007058 Code No : 03.QC.SC.09
	Enumeration of microorganisms - colony count technique at 30 OC	ISO 4833- 1:2013/Amd 1: 2022 Incubator Model : IOP 700 , , S.N: K709.0061 memmert – Germany Code No 03.QC.bf.73
	Enumeration of coliforms- colony count technique	ISO 4832:2006 Incubator Model: ICP 7000, S.N: K709.0062 memmert – Germany Code No: 03.QC.SC.72
	Horizontal method for detection of Salmonella Spp.	ISO 6579-1 :20 17/Arnd 2020 Incubator Model: BT –BTB20 – Egypt Code No : 03.QC.INC 126
	Enumeration of Yeasts & Moulds in products with water activity < 0.95	ISO 21527-2:2008 Incubator Model : Raypa ID- I50. S.N: 52455 UK
	Enumeration of Bacillus cereus	ISO 7932:2004/Amd 1:2020 Incubator Model: BST5020. S.N: 20189 Spain Code No 03.QC.BF.60
	Enumeration of Staphylococcus aureus	ISO 6888-1:2021/Amd 1:2023 Incubator Model M-780 , ser No. 18783 – Spain Code No 03.QC.BF10
	Enumeration of Entrobacteriaceae	ISO 21528-2:2017 Autoclave Model: LS-75HD , Ser No 23L-4926 UK Code No 03.QC.OT.154 Oven Model : STF-F12011 .S.N: R 282455 Spain Code No. 03.QC.OV.56

Kornish El Maadi, Riad El Maadi Tower # (1) - Cairo - Egypt

Tel.:(202)25275220/25275225/25275226/25275227

25240390/25249092

Fax: (202) 25275224

www.madein.eg منصة مصر الصناعية الرقمية

F4WI4TCL

1.1 / Jul 2024

Page 1 of 2

كورنيش المعادى – برج رياض المعادى رقم (1) - القاهرة – مصر

تليفون : 25275220/ 25275225 / 25275226 /25275227 (202)

25249092/ 25240390

فاكس : 25275224 (202)

الصفحة الرسمية للمجلس الوطني للاعتماد www.egac.gov.eg

Schedule of Accreditation
for Testing Laboratory According to ISO/IEC 17025
Issued To
(Quality Control Laboratory)
(Atos Pharma Company – Sekem)
(Elkatiba – Belbies City – Sharkeia Government - Egypt)

Schedule No.:020906B

1st Accreditation Date:24-Dec-2009

Issue No. (06):24-Dec 2025

Revision No. (0):

Valid to:23-Dec-2029

Materials / Products Tested	Types of Tests / Properties Measured / Range of Measurements	Standard Specifications / Techniques Used
Water	Determination of Fecal coliforms	SMWW 9222 D: 2023
	Determination of Total Plate Count at 35°C	SMWW 9215 B: 2023
	Determination of Total Coliform (MF)	SMWW 9222 B: 2023
	Determination of Fecal Streptococci (MF)	SMWW 9230 C: 2023
	Detection of Protozoa by microscopic examination by wet mount	SMWW 9711 A: 2023

Balance
Model: Vibra AJ 2200 E S.N: BLI41007058
Code No : 03.QC.SC.09
Incubator
Model : IOP 700 , , S.N: K709.0061
memmert – Germany Code No 03.QC.bf.73
Incubator
Model: ICP 7000, S.N: K709.0062
memmert – Germany Code No: 03.QC.SC.72
Incubator
Model: BT –BTB20 – Egypt
Code No : 03.QC.INC 126
Incubator
Model : Raypa ID- I50. S.N: 52455 UK
Incubator
Model: BST5020. S.N: 20189 Spain
Code No 03.QC.BF.60
Incubator
Model M-780 , ser No. 18783 – Spain
Code No 03.QC.BF10
Autoclave
Model: LS-75HD , Ser No 23L-4926 UK
Code No 03.QC.OT.154
Oven
Model : STF-F12011 .S.N: R 282455 Spain
Code No. 03.QC.OV.56

Kornish El Maadi, Riad El Maadi Tower # (1) - Cairo - Egypt

Tel.: (202) 25275220/25275225/25275226/25275227

25240390/25249092

Fax: (202) 25275224

www.madein.eg منصة مصر الصناعية الرقمية

F4WI4TCL

1.1 / Jul 2024

Page 2 of 2

كورنيش المعادي – برج رياض المعادي رقم (1) - القاهرة - مصر

تليفون : 25275220 / 25275225 / 25275226 / 25275227 (202)

25249092 / 25240390

فاكس : 25275224 (202)

الصفحة الرسمية للمجلس الوطني للاعتماد www.egac.gov.eg

Schedule of Accreditation
for Testing Laboratory According to ISO/IEC 17025
Issued To
(Quality Control Laboratory)
(Atos Pharma Company – Sekem)
(Elkatiba – Belbies City – Sharkeia Government - Egypt)

Schedule No.:020906B

1st Accreditation Date:24-Dec-2009

Issue No. (06):24-Dec 2025

Revision No. (0):

Valid to:23-Dec-2029

Materials / Products Tested	Types of Tests / Properties Measured / Range of Measurements		Standard Specifications / Techniques Used	
Water	Determination of elements	LOQ (mg/l)		
	Silica (SiO ₂)	0.4	SMWW 4500 Si02-C: 2023	Spectrophotometer Shimadzu – Japan, Model: UV 1650 PC S.N: A 10774501653
	Fluoride (F ⁻)	0.1	SMWW 4500 F-D: 2023	
	Sulphate (SO ₄)	12	SMWW 4500 S04-E: 2023	
	Iron (Fe)	0.2	SMWW 3500Fe ⁺³ -B: 2023	
	Determination of Chloride (Cl)		SMWW 4500Cl - B: 2023	Titrimetric Method Glass burette Code No. 03.QC.gb.92.1 (50ml) Code No. 03.QC.gb.92.2 (25ml)
	Determination of Calcium (Ca)		SMWW 3500Ca -B: 2023	
	Determination of Magnesium (Mg)		SMWW 3500Mg -B: 2023	
	Determination of Total Hardness		SMWW 2340C: 2023	
	Determination of Ca Hardness		SMWW 3500Ca-B: 2023	
	Determination of Mg Hardness		SMWW 3500Mg-B: 2023	
	Determination of Total alkalinity		SMWW 2320-B: 2023	
	Determination of Alkalinity as OH ⁻			
	Determination of Alkalinity as CO ₃			
Determination of Alkalinity as HCO ₃				

Seleen Gaber
19/11/2025

19-11-25

Kornish El Maadi, Riad El Maadi Tower # (1) - Cairo -Egypt

Tel.:(202)25275220/25275225/25275226/25275227

25240390/25249092

Fax: (202) 25275224

www.madein.eg منصة مصر الصناعية الرقمية

F4WI4TCL

1.1 / Jul 2024

Page 3 of 6

كورنيش المعادي – برج رياض المعادي رقم (1) - القاهرة – مصر

تليفون : 25275220 / 25275225 / 25275226 / 25275227 (202)

25249092/ 25240390

فاكس : 25275224 (202)

الصفحة الرسمية للمجلس الوطني للاعتماد www.egac.gov.eg

Schedule of Accreditation
for Testing Laboratory According to ISO/IEC 17025
Issued To
(Quality Control Laboratory)
(Atos Pharma Company – Sekem)
(Elkatiba – Belbies City – Sharkeia Government - Egypt)

Schedule No.:020906B

1st Accreditation Date:24-Dec-2009

Issue No. (06):24-Dec 2025

Revision No. (0):

Valid to:23-Dec-2029

Materials / Products Tested	Types of Tests / Properties Measured / Range of Measurements	Standard Specifications / Techniques Used	
Herbs & Spices	Determination of Essential oils in herbs	ISO 6571:2008/Amd 1:2017 (Modified) Eur.Ph 11/2023, Test No. 2.8.12	Balance Accuris Model W3200 series UK S.N. 1707644 Code No. 03.QC.SC.37
	Determination of acid insoluble ash	ISO 930: 1997 Eur.Ph 11/2023, Test No. 2.8.1	Balance Shimadzu Model ATX224R – Japan S.N. D320000253 Code no 03.QC.SC.20 Muffle Furnace Thermo Model F 6010 S.N 1249080398815 – UK Code No. 03.QC.OV.02
	Determination of total ash	ISO 928: 1997 Eur.Ph 11/2023, Test No. 2.4.16	Balance Shimadzu Model ATX224R – Japan S.N. D320000253 Code no 03.QC.SC.20 Muffle Furnace Thermo Model F 6010 S.N. 1249080398815 – UK Code No. 03.QC.OV.02
	Determination of loss on drying	Eur.Ph 11/2023, Test No. 2.2.:32 , USP test <921> method III (Gravimetric)	Balance Accuris Model W3200 series UK Sr No. 1707644 Code No. 03.QC.SC.37 Drying Oven

Kornish El Maadi, Riad El Maadi Tower # (1) - Cairo -Egypt

F4W14TCL

Tel.:(202)25275220/25275225/25275226/25275227

25240390/25249092

Fax: (202) 25275224

www.madein.eg منصة مصر الصناعية الرقمية

1.1 / Jul 2024

Page 4 of 6

A. P. Embady
A. ElKady 19/11/2025

كورنيش المعادي – برج رياض المعادي رقم (1) - القاهرة – مصر

تليفون : (202) 25275220/ 25275225 / 25275226 /25275227

25249092/ 25240390

فاكس : (202) 25275224

الصفحة الرسمية للمجلس الوطني للاعتماد www.egac.gov.eg

Schedule of Accreditation
for Testing Laboratory According to ISO/IEC 17025
Issued To
(Quality Control Laboratory)
(Atos Pharma Company – Sekem)
(Elkatiba – Belbies City – Sharkeia Government - Egypt)

Schedule No.:020906B

1st Accreditation Date:24-Dec-2009

Issue No. (06):24-Dec 2025

Revision No. (0):

Valid to:23-Dec-2029

Materials / Products Tested	Types of Tests / Properties Measured / Range of Measurements	Standard Specifications / Techniques Used	
Herbs & Spices	Determination of bulk density	ASTA method 25: 0 Eur. Ph 11, 2023, Method 2.9.34.1.	Balance Kern Model 3600-2N series no W2006975 Code no 03.QC.SC.08
	Estimation of particle size distribution	Eur. Ph 11 Method 2.9.38: 2023	Balance Accuris Model W3200 series UK Sr No. 1707644 Code No. 03.QC.SC.37 Sieve Shaker – Germany Retsch, S.N 83004022 2.00 mm S.N 461909 1.00 mm S.N 437306 630 µm S.N 319041 280 µm S.N 316794 200 µm S.N 380283-
Honey	Determination of reducing sugar	ES 355-2: 2005 CODEX:1987	Balance Accuris Model W3200 series UK Sr No. 1707644 Code No. 03.QC.SC.37 Glass burette Code No. 03.QC.gb.92.1 (50ml) Code No. 03.QC.gb.92.2 (25ml)
	Determination of sucrose content		

Kornish El Maadi, Riad El Maadi Tower # (1) - Cairo -Egypt

Tel.:(202)25275220/25275225/25275226/25275227

25240390/25249092

Fax: (202) 25275224

www.madein.eg منصة مصر الصناعية الرقمية

F4WI4TCL

1.1 / Jul 2024

Page 5 of 6

M.P. Enbab
A. ElKady 19/11/2025

OK
19.11.25

كورنيش المعادي – برج رياض المعادي رقم (1) - القاهرة – مصر

تليفون : 25275220/ 25275225 / 25275226 /25275227 (202)

25249092/ 25240390

فاكس : 25275224 (202)

الصفحة الرسمية للمجلس الوطني للاعتماد www.egac.gov.eg

Schedule of Accreditation
for Testing Laboratory According to ISO/IEC 17025
Issued To
(Quality Control Laboratory)
(Atos Pharma Company – Sekem)
(Elkatiba – Belbies City – Sharkeia Government - Egypt)

Schedule No.:020906B

1st Accreditation Date:24-Dec-2009

Issue No. (06):24-Dec 2025

Revision No. (0):

Valid to:23-Dec-2029

Materials / Products Tested	Types of Tests / Properties Measured / Range of Measurements	Standard Specifications / Techniques Used	
Pharmaceutical products	Tomex Tablets : Determination of allicin	In-House Method 0009 based on Eur.P.h 11	High Performance Liquid Chromatography (HPLC) Model 1100 series S.N: DE43634075 Code No. 03.QC.hp.64 Balance Shimadzu Model ATX224R – Japan Code no 03.QC.SC.20
	Tomex Plus tablets : Determination of allicin	In-House Method 0010 based on Eur.Ph 11	
	Neuropatex SGC : Determination of alpha lipoic acid	In-House Method 0023 based on USP:2025	
	Thymotal Syrup : Determination of thymol	In-House Method 0059	
	Cystinol HGC : Determination of total flavonoids as Rutin	In-House Method 0022	Balance Shimadzu Model ATX224R – Japan Code no 03.QC.SC.20 Spectrophotometer Model: UV 1650 PC S.N: A 10774501653 Shimadzu – Japan Code No. 03.QC.SP.28

AR d Hajez 19-11-2025

19-11-25

Kornish El Maadi, Riad El Maadi Tower # (1) - Cairo -Egypt

Tel.:(202)25275220/25275225/25275226/25275227

25240390/25249092

Fax: (202) 25275224

www.madein.eg منصة مصر الصناعية الرقمية

F4WI4TCL

1.1 / Jul 2024

Page 6 of 6

كورتيش المعادى – برج رياض المعادى رقم (1) - القاهرة – مصر

تليفون: ٢٥٢٧٥٢٢٠ / ٢٥٢٧٥٢٢٥ / ٢٥٢٧٥٢٢٦ / ٢٥٢٧٥٢٢٧ (٢٠٢)

٢٥٢٤٩٠٩٢ / ٢٥٢٤٠٣٩٠

فاكس: ٢٥٢٧٥٢٢٤ (٢٠٢)

الصفحة الرسمية للمجلس الوطني للاعتماد www.egac.gov.eg